



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

### PUNE

#### CLARIFICATION ON TENDER NUMBER - IISER-PUR-0356-17

ITEM DESCRIPTION- PROCUREMENT OF DATA ACQUISITION SYSTEM AND BROADBAND SEISMOMETER. Refer our Press Tender Notice No.IISER/S&P/13/17 for procurement of DATA ACQUISITION SYSTEM AND BROADBAND SEISMOMETER. Tender Reference Number - IISER-PUR-0356-17.

Pre-Bid meeting was held on June 28<sup>th</sup> , 2017 at 2.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

28.6.2017

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR DATA ACQUISITION SYSTEM AND BROADBAND SEISMOMETER**

## TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0356-17

DATE : 28.6.17

Sr.No	Query/Clarification Sought	Clarification / Amendment
01	Whether One Recorder with three independent channels?	One recorder with three independent 24-bit digitisers, one for each channel
02	Since this is being used with Broadband system, why you need event trigger mode? Event Trigger used for Strong Motion System. If you want to derive the triggered events, you can have through the Processing Software.	No change in tender specifications
03	Why SPS 1? It should be atleast 50 and above. Because user need to use atleast 100sps or more.	No change in tender specifications

04	Should be GPS based Time. Globally all are using GPS based timing.	No change in tender specifications
05	<ul style="list-style-type: none"> <li>i. Calibration facility for Broadband seismometer from DAS</li> <li>ii. Mass position monitoring for Broadband seismometer <u>or auto.</u></li> </ul> Mass centering on command for Broadband seismometer <u>or auto mass centering.</u>	No change in tender specifications
06	<ul style="list-style-type: none"> <li>i. RAM 16 MB or more Q: RAM is old concept. Now-a-days system does not require RAM. CF Card is used for memory. RAM does not improve in data quality or recording.</li> <li>ii. User removable Recording media of capacity 32 GB or more</li> <li>iii. Hot swappable recording media Q: in above it is mentioned CF Card removable.</li> </ul>	No change in tender specifications
07	Mass position monitoring from the data acquisition system <u>Or, Auto.</u> Mass centering on command from the data acquisition system <u>or, auto mass centering</u>	No change in tender specifications
08	For portable application, 3 - 5m antenna cable length is sufficient. Please comment, is this for portable or observatory installation.	No change in tender specifications
09	We do have built-in storage capacity of 8GB as standard; we believe that additional 8GB capacity of removable storage is sufficient - please confirm.	No change in tender specifications

10	Please comment, do we need to supply the VSAT communication equipment for data transmission?	No
11	We do have built-in LED indicators for verifying the system operational status. Further the data viewing, monitoring can be done through WIFI enabled mobile tablet / laptop. We believe that the external monitor may not be required in such case, please confirm.	No change in tender specifications



## IISER PUNE

## PRE-BID CONFERENCE FOR DATA ACQUISITION SYSTEM AND BROADBAND SEISMOMETER

## COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0356-17

DATE : 28.6.17

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>As per Page# 5, Technical Bid Point# h) Solvency certificates (not older than twelve months) issued by Scheduled / Nationalized bank with which BIDDER holds the current account.</p> <p><b>Query:</b> In case of foreign bidder, whether the letter regarding solvency of the bidder from a foreign bank where the bidder holds the accounts is acceptable?</p>	<p><b>Page# 5, Technical Bid Point# h- Solvency Certificate</b></p> <p>In case of foreign bidder, the letter regarding solvency of the bidder from a foreign bank where the bidder holds the accounts is acceptable. However, it should be routed through Indian Bank.</p>
2	<p>As per Page# 10, 2. Delivery Period / Timeliness: The delivery &amp; installation must be completed within 03 months after the placement of Purchase Order.</p> <p><b>Query:</b> We request you to amend it as "The delivery &amp; installation must be completed within 03 months after the placement of</p>	<p><b>As per Page# 10, 2. Delivery Period / Timeliness - Delivery Period / Timeliness clause is amended as:</b></p> <p>The delivery &amp; installation must be completed within 03 months after the opening of Letter of Credit.</p>

	<b>Purchase Order and opening of Irrevocable Letter of Credit".</b>	
3	<p><b>As per Page# 10, 3. Security Deposit:</b> Within ten (10) days of the award of contract, the vendor shall furnish a Security Deposit amounting to 10% of the Purchase Order value in the form of Demand Draft / Bank Guarantee (from scheduled bank only) favoring the Director, Indian Institute of Science Education and Research, Pune.</p> <p>Query: We request you to amend as "Within fifteen (15) days of the award of contract, the vendor shall furnish a Security Deposit amounting to 10% of the Purchase Order value in the form of Demand Draft / Bank Guarantee (from scheduled bank only) favoring the Director, Indian Institute of Science Education and Research, Pune".</p>	<p><b>As per Page# 10, 3. Security Deposit clause is amended as:</b></p> <p>Within fifteen (15) days of the award of contract, the vendor shall furnish a Security Deposit amounting to 10% of the Purchase Order value in the form of Demand Draft / Bank Guarantee (from scheduled bank only) favoring the Director, Indian Institute of Science Education and Research, Pune.</p>
4	<p><b>As per discussion:</b></p> <p>i) The installation &amp; training will be held at IISER, Pune. ii) No abroad training is required.</p>	<p><b>Installation &amp; Training</b></p> <p>Yes, the installation &amp; training will be held at IISER Pune and No abroad training is required.</p>
5	<p><b>As per Page# 13, 8. Inspection, 8.3:</b> The supplier should arrange for physical Inspection of the items directly or through their authorized representative within seven days of arrival of the consignment failing which they will be responsible for the losses. After the shipment is effected, the supplier / its representative / India agents must remain in touch with the lab / instt. to ascertain the date of arrival of consignment.</p> <p>Query: We request you to amend as "The</p>	<p><b>Page# 13, 8.3 Inspection Clause is amended as:</b></p> <p>"The supplier should arrange for physical Inspection of the items directly or through their authorized representative within fifteen days of arrival of the consignment failing which they will be responsible for the losses. After the shipment is effected, the supplier / its representative / India agents must remain in touch with the lab / instt. to ascertain the date of arrival of consignment".</p>

<p>supplier should arrange for physical Inspection of the items directly or through their authorized representative within fifteen days of arrival of the consignment failing which they will be responsible for the losses. After the shipment is effected, the supplier / its representative / India agents must remain in touch with the lab / instt. to ascertain the date of arrival of consignment" because as per tender the installation should be completed within 15 days from the date of arrival at IISER, Pune if you extend the inspection period as 15 days within the same visit we shall complete both Inspection, Installation &amp; Training.</p>	
--	--