

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1160-14

ITEM DESCRIPTION- PROCUREMENT OF PROTON CONDUCTIVITY MEASUREMENT SYSTEM

Refer our Press Tender Notice No.IISER/S&P/10/14 dated 10.12.2014 for procurement of Proton Conductivity Measurement System. Tender Reference Number - IISER-PUR-1160-14.

Pre-Bid meeting was held on December 23, 2014 at 3.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac .in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

DATE: 23.12.14



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF PROTON CONDUCTIVITY MEASUREMENT SYSTEM

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1160-14

S.No	Query/Clarification Sought	Clarification / Amendment	
1	Page 23: point number 1 (i) 7th line from the bottom Whether Instrument should have provision to operate easily through only front panel.	Page 23: point number 1 (i) 7th line from the bottom We have mentioned "Instrument should have provision to operate easily through front panel and also using computer using specialized software." Front panel is not important but it should have provision to operate easily. has been modified as: "Instrument should have provision to operate easily and also using computer using specialized software."	
2	Page 23: point number 1 (iii) and Page 24: point number 2 first two line Whether Temperature and Humidity chamber both should be controllable through impedance analyzer software.	Page 23: point number 1 (iii) and Page 24: point number 2 (first two line)- Separate software is also fine. has been modified as- Temperature and Humidity chamber should be controllable through software	

DATE: 23.12.14



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF PROTON CONDUCTIVITY MEASUREMENT SYSTEM

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1160-14

S.No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL