

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0634-16

ITEM DESCRIPTION- PROCUREMENT OF ELECTROCHEMICAL WORK STATION

Refer our Press Tender Notice No.IISER/S&P/8/16 dated 20.9.2016 for procurement of Electrochemical Work Station. Tender Reference Number - IISER-PUR-0634-16.

Pre-Bid meeting was held on September 29th, 2016 at 3.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac .in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

Sd/-Assistant Registrar (S&P)

DATE: 29.9.16



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ELECTROCHEMICAL WORK STATION

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0634-16

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No - 25, Chapter 4 Clarification regarding "UV-Vis-NIR wavelength range 200-2500 nm"	Tender Specification prevails. No change in the specification. Control and data acquisition for both UV-Vis and Vis-NIR should be done simultaneously, resulting in single data plot only.
2	Page No -24, Chapter 4 Clarification regarding "Potential Rise/Fall Time"	Tender Specification prevails. No change in the specification. Should be minimum 300 ns or lower

3	Page No - 24 , Chapter 4 Clarification regarding "Compliance Voltage: ±18 V or better at ±350 mA"	Tender Specification prevails. No change in the specification. Compliance Voltage higher than ±18V is acceptable.
4	Page No - 24, Chapter - 4 Clarification regarding "Signal generator frequency range 10µHz-25MHz"	Tender Specification prevails. No change in the specification. We have specified higher frequency range of signal generator so that the system remains modular towards higher frequency Electrochemical Impedance Spectroscopy measurements using suitable device for future applications.



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ELECTROCHEMICAL WORK STATION

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0634-16 DATE : 29.9.16

S.No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL