भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपती

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar, Karakambadi Road, Mangalam (B.O), Tirupati - 517 507



MENTORED BY

Indian Institute of Science Education and Research Pune Dr. Homi Bhabha Raod, Pune: 411 008 (Maharashtra) www.iiserpune.ac.in

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0128-15

ITEM DESCRIPTION- PROCUREMENT OF ANALYTICAL CUM SEMIPREPARATIVE HPLC SYSTEM

Refer our Press Tender Notice No.IISER/S&P/08/15-16 dated 24.11.2015 for procurement of Analytical Cum Semipreparative HPLC System. Tender Reference Number - IISERT-PUR-0128-15.

Pre-Bid meeting was held on December 04th, 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac .in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

DATE: 4.12.2015



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ANALYTICAL CUM SEMIPREPARATIVE HPLC SYSTEM

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0128-15

S.No	Query/Clarification Sought	Clarification / Amendment
1	Solvent Delivery System for Semi-prep & Analytical flow rates- Page No - 25, Chapter 4, Point-1(2) Suggestion by one of the bidder "Pump pressure drops if the flow rate is above 5 ml/minute. Hence we request you to change this point as "Pressure range 5000psi upto 5ml flow rate"	Constant-pressure deliver possibility throughout the range is a crucial feature in terms of semi preparative operations. Lowering this range to 5000psi up to 5 ml/minute is not possible. Hence, tender Specification prevails. No change in the specification.

2	Page No - 26, Chapter 4, Point - 4(8)	Tender Specification clearly states "up to 80 Hz",
	Data Acquisition: Up to 80 Hz	hence 20 Hz will technically comply.
	Suggestion by one of the bidder: In PDA / DAD, 80 Hz data acquisition is required only when very sensitive analytical work is conducted in HPLC system. In semi preparative HPLC normally standard data acquisition 20 Hz is used. We request to amend it to 20 Hz.	



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ANALYTICAL CUM SEMIPREPARATIVE HPLC SYSTEM

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0128-15 DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL