



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1113-13

ITEM DESCRIPTION- PROCUREMENT OF HIGH PERFORMANCE LIQUID CHROMATOGRAPHY SYSTEM WITH BOTH UV AND ELSD DETECTOR

Please refer our Press Tender Notice No.IISER/S&P/13/13 dated 27.11.2013 for procurement of High Performance Liquid Chromatography System With Both UV And ELSD Detector . Tender Reference Number - IISER-PUR-1113-13.

Pre-Bid meeting was held on 09th December 2013 at 11.00 AM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair.

9.12.2013

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE LIQUID CHROMATOGRAPHY SYSTEM WITH BOTH UV AND ELSD DETECTOR

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER -IISER-PUR-1113-13

DATE : 9.12.13

S.No	Query/Clarification Sought	Clarification / Amendment
1	Clarification on the requirement of columns.	Columns will be ordered separately as per requirement.
2	Chapter 4, Page No. 24 Why are 1024 elements or more needed for diode array detector?	Chapter 4, Page No. 24 To increase the sensitivity and resolution of the data. Our applications will involve analysis of small quantities of biological samples which are very similar structurally and their resolution is critical for our experiments. The tender specifications remain unchanged.
3	Chapter 4, Page No. 24 Why is the required wavelength range for detector 190-950 nm and not upto 800 nm?	Chapter 4, Page No. 24 Detection upto 950 nm is required for applications that involve purification of near IR (infra-red) dyes and compounds that absorb at that region of the spectrum. These applications will be critical for our research and so having the detector detect compounds upto 950 nm is critical. The tender specifications remain unchanged.



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE LIQUID CHROMATOGRAPHY
SYSTEM WITH BOTH UV AND ELSD DETECTOR**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER -IISER-PUR-1113-13

DATE : 9.12.13

S.No	Query/Clarification Sought	Clarification / Amendment
1	-----NIL -----	-----NIL-----