



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0062-18

ITEM DESCRIPTION- PROCUREMENT OF SCANNING ELECTROCHEMICAL WORKSTATION WITH IMPEDANCE MODE

Refer our Press Tender Notice No.IISER/S&P/2/2018 dated 5.5.2018 for procurement of Scanning Electrochemical workstation with impedance mode.Tender Reference Number - IISER-PUR-0062-18.

Pre-Bid meeting was held on May 10th , 2018 at 2.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

10.5.2018

Sd/-
Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF SCANNING ELECTROCHEMICAL WORKSTATION
WITH IMPEDANCE MODE**

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0062-18

DATE :10.5.18

S.No	Query/Clarification Sought	Clarification / Amendment
1	Chapter 4; Page 24 Impedance: 10 T Ohms or better	Chapter 4; Page 24 It is changed to; Impedance: 1 T Ohms or better

2	<p>Chapter 4; Page 24</p> <p>Frequency: 1 μHz-1 MHz or wider</p>	<p>Chapter 4; Page 24</p> <p>It is changed to;</p> <p>Frequency: 10 μHz-1 MHz or wider</p>
3	<p>Chapter 4; Page 25</p> <p>Scan range (x,y,z) : 100 mm or greater</p>	<p>Chapter 4; Page 25</p> <p>Tender specifications remain unchanged</p>
4	<p>Chapter 4; Page 27</p> <p>Crystal: 9 MHz AT-cut: Gold as well as Platinum</p>	<p>Chapter 4; Page 27</p> <p>Tender specifications remain unchanged</p>