



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0121-16

ITEM DESCRIPTION- PROCUREMENT OF ELECTROCHEMICAL WORKSTATION

Refer our Press Tender Notice No.IISER/S&P/02/2016 dated 21.6.2016 for procurement of Electrochemical Work Station.Tender Reference Number - IISER-PUR-0121-16.

Pre-Bid meeting was held on July 01st , 2016 at 11.00 AM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

1.7.2016

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ELECTROCHEMICAL WORKSTATION

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0121-16

DATE:01.07.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No -25, Chapter 4, Clarification regarding "number of channels can be updated"	The system should be upgradable upto 16 channels
2	Page No -25, Chapter 4 Regarding voltage range of Channels 1,2 and 3	Amendment in Page No - 25, Chapter 4 Voltage range for channels 1 2 and 3 should be $10 V \leq V \leq -10V$
3	Page No - 25 , Chapter 4 Regarding Bandwidth	Amendment in Page No - 25, Chapter 4 Bandwidth $\geq 1\text{MHz}$



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ELECTROCHEMICAL WORKSTATION

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0121-16

DATE : 1.7.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----