



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0181-15

ITEM DESCRIPTION- PROCUREMENT OF CIRCULAR DICHROISM SPECTROMETER

Refer our Press Tender Notice No.IISER/S&P/20/15-16 dated 1.3.2016 for procurement of Circular Dichroism Spectrometer . Tender Reference Number - IISER-PUR-0181-14.

Pre-Bid meeting was held on March 09<sup>th</sup> , 2016 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

9.3.2016

Sd/-  
Assistant Registrar (S&P)



IISER PUNE

## PRE-BID CONFERENCE FOR PROCUREMENT OF CIRCULAR DICHROISM SPECTROMETER

### TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0181-15

DATE : 9.3.2016

S.No	Query/Clarification Sought	Clarification / Amendment
1.	<p>Page No - 25, Chapter 4, Point no. 3</p> <p>The instrument must have double prism monochromator and a low stray light</p> <p>&lt;0.0003 % at 200 nm</p> <p>Suggestion by one of the bidders "double grating should also work in the same way"</p>	<p>Page No - 25, Chapter 4, Point no. 3</p> <p>The instrument must have double prism or double grating monochromator and a low stray light &lt;0.0003 % at 200 nm</p>

2.	<p>Page No - 25, Chapter 4, Point no. 8</p> <p>CD resolution must be 0.005-0.008 mdeg(at ±10 mdeg full scale)</p> <p>Suggestion by one of the bidders "Please ask for CD resolution of 0.00001mdeg over CD scale of + 8000 mdeg with automatic scaling"</p>	Tender specification prevails
3.	<p>Page No -25, Chapter 4, Point no. 9</p> <p>Baseline stability should be 0.03 mdeg/hr or better</p> <p>Suggestion by one of the bidders"Baseline stability should be specified as mdeg/ ° C not mdeg/hr as it is temperature dependent"</p>	Tender Specification prevails.
4.	<p>Page No - 25 , Chapter 4 , point no 10</p> <p>RMS noise at 185 nm must be &lt;0.5 mdeg</p> <p>RMS noise at all wavelengths including 175 nm with Data integration time</p>	Tender specification prevails
5.	<p>Page No - 26 , Chapter 4 , point no 23</p> <p>LD capability should be an optional accessory or main</p>	LD feature is very important for our operations and hence should be offered in the main quote itself
6.	<p>Page No - 26 , Chapter 4 , point no 24</p> <p>low volume capillary instead of cuvettes</p>	Tender specification prevails



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF CIRCULAR DICHROISM SPECTROMETER****COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISER-PUR-0181-15

DATE : 9.3.2016

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----