



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

ADDENDUM ON TENDER NUMBER - IISER-PUR-1491-17

ITEM DESCRIPTION- PROCUREMENT OF HIGH PERFORMANCE LIQUID CHROMATOGRAPHY SYSTEM WITH ACCESSORIES

Refer our Press Tender Notice No.IISER/S&P/1491/17 dated 10.3.2018 for procurement of High Performance Liquid Chromatography System with accessories and minutes of Pre-Bid Conference dated 15th March 2018. Tender Reference Number - IISER-PUR-1431-17.

This is in continuation to our Minutes of Pre-Bid Conference dated 15.3.2018. An addendum on Technical queries clarification has been made on minutes of Pre-Bid Conference. The same is attached as Annexure - IV

All the Prospective Bidders are required to take cognizance of the addendum along with proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged.No more correspondence in this regard will be entertained

3.4.2018

Sd/-
Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE LIQUID
CHROMATOGRAPHY SYSTEM WITH ACCESSORIES**

ADDENDUM ON CLARIFICATION

TENDER NUMBER - IISER-PUR-1491-17

DATE: 3.4.18

S.No	Query/Clarification Sought	Clarification / Amendment
18	<p data-bbox="184 857 495 889">Chapter 4, section 9</p> <p data-bbox="184 967 1199 1068">We request you to please clarify for what applications exactly are you requiring Chiral columns? Please specify more information than the dimensions mentioned</p>	<p data-bbox="1251 857 1562 889">Chapter 4, section 9</p> <p data-bbox="1251 967 1969 1068">We want reverse phase semi preparative HPLC column (5 U, C18(2), 100A, 250*10mm) instead of chiral column.</p>