



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-2215-20

ITEM DESCRIPTION- PROCUREMENT OF OPTICAL 3D SURFACE PROFILOMETER.

Refer IISER Pune tender notice dated 19/04/2021 for procurement of Optical 3D surface profilometer.

Pre-Bid meeting was held on April 27<sup>th</sup>, 2021 at 4.00 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

Date: 27/04/2021

Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF Optical 3D surface profilometer**

**TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION**

| S.No | Query/Clarification Sought  | Clarification / Amendment   |
|------|---|---|
| 1    | Our instrument is having 4 turret FOV lens holder as standard. You have opted for 5 turret, can you specify as 4-5 turret   | Single Objective Mount with an option to upgrade to 4-5 or more objectives manual turret or automated turret must be available. |
| 2    | Diameter of an atom falls in the range of 0.1 -0.5nm, you have specified measurement range and repeatability as 0.1nm & 0.01nm, we assume the numbers are theoretically better numbers. We request you to specify as resolution 1nm, repeatability 0.1nm so that specifications are more realistic. | No change. Resolution and repeatability remains same as mentioned in the tender. This is required for our experiment.           |
| 3    | Since many of your optional features are standard in our instrument can you increase the budget by Rs 5 lakhs   | No change.  |