



# INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0218-16

ITEM DESCRIPTION- PROCUREMENT OF STAND ALONE ZEBRAFISH  
SYSTEM

Refer our Press Tender Notice No.IISER/S&P/02/2016 dated 21.6.2016 for procurement of Stand Alone Zebra System.Tender Reference Number - IISER-PUR-0218-16.

Pre-Bid meeting was held on June 30<sup>th</sup>, 2016 at 4.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

30.06.2016

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF STAND ALONE ZEBRAFISH SYSTEM**

## TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0218-16

DATE : 30.6.14

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>Technical specification point 3</p> <p>A bidder has requested clarification if the requirement for self-cleaning technology for filtration can be dropped.</p>	<p>We need filter-free technology.</p> <p>Tender Specification prevails.</p> <p>No change in the specification.</p>
2	<p>Technical specification point 23</p> <p>A bidder has requested clarification whether self-cleaning tanks qualify under this point.</p>	<p>Yes, self cleaning tank technology will qualify. Point 23 is amended to:</p> <p>Each tank should have a continuous tank siphoning system by means of a siphon are similar technology that allows for self cleaning of tanks.</p>



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF STAND ALONE ZEBRAFISH SYSTEM  
SYSTEM**

## COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0218-16

DATE : 30.6.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----