



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0987-16

ITEM DESCRIPTION- PROCUREMENT OF HELIUM LIQUEFIER

Refer our Press Tender Notice No.IISER/S&P/18/16 dated 26.12.2016 for procurement of Helium Liquefier. Tender Reference Number - IISER-PUR-0987-16.

Pre-Bid meeting was held on January 03rd , 2016 at 2.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

3.1.2017

Sd/-
Assistant Registrar (S&P)



IISER PUNE
PRE-BID CONFERENCE FOR PROCUREMENT OF HELIUM LIQUEFIER
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0987-16

DATE : 3.1.2017

S.No	Query/Clarification Sought	Clarification / Amendment
1	Tender includes specification for Helium Liquefier, Purification and Storage. But there is no mention of Helium recovery system part. According to us with Helium Recovery system just a Helium Liquefier and Purifier and storage will not be helpful and complete. So, please give us broad specifications of Helium Recovery system and whether this needs to be quoted / offered.	IISER Pune would be installing a Helium recovery system on site. However, this would be done separately, and is beyond the scope of this tender. Hence the vendor need not quote for the Helium recovery system.



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF HELIUM LIQUEFIER**COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISER-PUR-0987-16

DATE : 3.1.2017

S.No	Query/Clarification Sought	Clarification / Amendment
1.	Can we quote three year warranty with separate prices and can be chosen and added and gives flexibility. While we are willing to give 3 year warranty please consider PBG of only one year.	Warranty prices can be shown separately. Accordingly PBG should be submitted for three years not one year for three years of warranty period.