

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0023-18

ITEM DESCRIPTION- PROCUREMENT OF GAS CHROMATOGRAPHY – TRIPLE QUARDRUPOLE MASS SPECTROMETRY SYSTEM

Refer our Press Tender Notice No.IISER/S&P/1/18 dated 26.4.2018 for procurement of Gas Chromatography - Triple Quardrupole Mass Spectrometry System. Tender Reference Number - IISER-PUR-0023-18.

Pre-Bid meeting was held on May 2, 2018 at 3.30 pm and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac .in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

Sd/-Assistant Registrar (S&P)

2.5.2018



IISER PUNE PRE-BID CONFERENCE FOR PROCUREMENT OF GAS CHROMATOGRAPHY - TRIPLE QUARDRUPOLE MASS SPECTROMETRY SYSTEM

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0023-18

DATE : 2.5.18

S.No	Query/Clarification Sought	Clarification / Amendment
1	In Chapter 4, page 25, point 13 in Gas Chromatograph,	Chapter 4, page 25, point 13
	tender mentions provision to install two or more columns in	
	the oven: 'At least two >100m capillary columns or two 20ft	Tender Specification prevails. No change in the specification.
	steel-packed columns.' Please change 'two >100m capillary	
	columns' to 'two 60m capillary columns'.	
2	In Chapter 4, page 26, point 3 in Triple quadrupole mass	Chapter 4, page 26, point 3 is amended as
	spectrometer, tender mentions that "The triple quadrupole	
	should be inert. It should be heatable (\geq 180°C) for self-	"The triple quadrupole should be inert. Quadrupole should
	cleaning". Please specify which quadrupole should be	be heatable (≥180°C) for self-cleaning"
	heatable.	