



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1413-17

ITEM DESCRIPTION- PROCUREMENT OF MICRO INJECTOR AND
MICROMANIPULATOR SYSTEM

Refer our Press Tender Notice No.IISER/S&P/30/17 dated 10.3.2018 for procurement of Micro Injector and Micromanipulator System. Tender Reference Number - IISER-PUR-1413-17.

Pre-Bid meeting was held on March 15th , 2018 at 3.30 pm and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

15.3.2018

Sd/-
Assistant Registrar (S&P)



**IISER PUNE
PRE-BID CONFERENCE FOR PROCUREMENT OF MICRO INJECTOR AND MICROMANIPULATOR
SYSTEM**

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1413-17

DATE : 15.3.18

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>Page No -18, Chapter 4</p> <p>Unit of pressure KPA (KILO PASCAL) should be accepted instead of HPA (Hectopascal)</p>	<p>Both kPa and hPa units can be accepted.</p>
2	<p>Page No -18, Chapter 4</p> <p>In specification sheet, For Injector (Injection Side): Everywhere, it is mentioned 87PSI as pressure & we have up to 60psi which is sufficient enough for all the application mentioned in the tender document.</p>	<p>For the Microinjector: the specifications are amended to :</p> <p>10. Injection pressure: 0; 60 psi, or better.</p> <p>11. Compensation pressure: 0; 60 psi, or better.</p> <p>12. Clean function: Maximum rinsing pressure of 60 psi, or better.</p> <p>13. Continuous flow function: 0; 60 psi; can be set in increments of 0.01 psi, or better.</p>