



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH  
PUNE

CLARIFICATION ON TENDER NUMBER - IISER/PUR/0554/20

ITEM DESCRIPTION- Procurement of Superconducting Magnet system with liquid helium cryostat

Refer IISER Pune open tender number IISER/PUR/0554/20 for procurement of Superconducting Magnet system with liquid helium cryostat.

**It is notified that, Tender is invited from manufacturers/local suppliers for above item manufactured in India strictly as per PPP-MII Order dated 4th June 2020 from all over India and INR payment terms only. NON LOCAL SUPPLIERS NEED NOT APPLY.**

Pre-Bid meeting was held on December 18<sup>th</sup>, 2020 at 4.00 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -I

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

18.12.2020

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF SUPERCONDUCTING MAGNET SYSTEM WITH LIQUID  
HELIUM CRYOSTAT  
TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION**

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p><b>Eligibility criteria -</b>  <b>Category of Suppliers invited for this Tender</b>            Class I local Supplier - has local content equal to more than 50%            Class II local Supplier - has local content more than 20% but less than 50%            Non -Local Supplier - has local content less than or equal to 20%</p>	<p><b>Eligibility criteria is amended to</b>   <u><b>Category of Suppliers invited for this Tender</b></u>            Class I local Supplier - has local content equal to more than 50%.            Class II local Supplier - has local content more than 20% but less than 50%.  <b>As per PPP-MII Order dated 4th June 2020 from all over India and INR payment terms only. NON LOCAL SUPPLIERS NEED NOT APPLY</b></p>
2	<p>Nominal Helium Capacity: A query was raised on whether this can be increased from the current value of 25 litres or more mentioned in the tender document.</p>	<p>This value quoted is the bare minimum, and the vendors are free to quote for any cryostat which meets (or exceeds) this capacity. Hence <b>No change in the tendered specifications</b></p>